S-Tec GmbH
Gewerbepark 13
87640 Biessenhofen
Germany

Certificate of Calibration

Cert No.: E0101-424_1
Project Number: P240057300

Customer:
Infineon Technologies Austria AG
Siemensstraße 2 Villach 9500
Austria

Asset ID: (IFAT)	232925	As Found:	**FAIL**
Equipment Type:	Thermostream	Adjusted:	Yes
Manufacturer:	Thermonics	As Left:	**PASS**
Model Number:	T2500-E	Cal. Date:	2025-03-26
Serial Number:	E0101-424	Cal. Interval:	12 months
Software version:		Due Date:	2026-03-26
Calibration Location:	Customer Facility	Climate Condition:	(21,5 bis 21,7) °C ±1 K (46 bis 47) % ±3 % (1010 hPa - 1017 hPa) hPa ±5 hPa
		Performed by:	M.Schumacher

Used Standards

Asset ID	Description	Model	Serial	Manufacturer	Due Date	Tracebility Number
161314-06	Kalibrator	714B	33010008	Fluke	Mai 25	

Used Procedures

Procedure Name (Revision Number)	Description
	According to manufacturer's instructions

Remark:

Approved by:

Schumacher

S-Tec GmbH
Gewerbepark 13
87640 Biessenhofen
Germany

Certificate of Calibration

2	Nozzle sensor type K Specification [± 2°C]					
Set Temperature [°C]	As Found Temperature [°C]	As Found Deviation	As Found Result	As Left Temperature [°C]	As Left Deviation	As Left Result
-40	-42,6	-2,60	FAIL	-40,4	-0,40	PASS
-20	-23,3	-3,30	FAIL	-20,5	-0,50	PASS
+25	+25,6	0,60	PASS	+25,4	0,40	PASS
+70	+74,4	4,40	FAIL	+70,5	0,50	PASS
+85	+87,5	2,50	FAIL	+85,2	0,20	PASS
+125	+130,2	5,20	FAIL	+125,3	0,30	PASS
+150	+155,3	5,30	FAIL	+150,6	0,60	PASS
+170	+175,6	5,60	FAIL	+170,3	-0,30	PASS

External sensor type K Specification [± 2°C]						
Set Temperature [°C]	As Found Temperature [°C]	As Found Deviation	As Found Result	As Left Temperature [°C]	As Left Deviation	As Left Result
-40	-40,2	-0,20	PASS			
-20	-20,4	-0,40	PASS			
+25	+25,2	0,20	PASS			
+70	+70,4	0,40	PASS			
+85	+85,1	0,10	PASS			
+125	+125,3	0,30	PASS			
+150	+150,3	0,30	PASS			
+170	+170,3	0,30	PASS			

If “As Found” Calibration **OK**, then only values in **“As Found”**
If optimized, then values in “As Found” and “As Left” **-> Remark required**
If adjusted (As Found FAIL), then values in “As Found” and “As Left” **-> Remark required**