

Certificate of Calibration

Cert No.: 240041304_1

Customer:

Infineon Technologies Austria AG
 Siemensstraße 2 Villach 9500
 Austria

Project Number:

Asset ID: (IFAT)		As Found:	PASS
Equipment Type:	Thermostream	Adjusted:	No
Manufacturer:	S-Tec	As Left:	PASS
Model Number:	TSR2252LN MH	Cal. Date:	2024-08-08
Serial Number:	240041304	Cal. Interval:	12 months
Software version:		Due Date:	2025-08-08
Calibration Location:	Manufacturer's Facility	Climate Condition:	(21,5 bis 21,7) °C ±1 K (46 bis 47) % ±3 % (1010 hPa - 1017 hPa) hPa ±5 hPa
		Performed by:	M.Schumacher

Used Standards

Asset ID	Description	Model	Serial	Manufacturer	Due Date	Traceability Number
161314-06	Kalibrator	714B	33010008	Fluke	Mai 25	

Used Procedures

Procedure Name (Revision Number)	Description
	According to manufacturer's instructions

Remark:

Approved by:

Schumacher

 Schumacher

Certificate of Calibration

2 Nozzle sensor type K Specification [± 2°C]						
Set Temperature [°C]	As Found Temperature [°C]	As Found Deviation	As Found Result	As Left Temperature [°C]	As Left Deviation	As Left Result
-40	-40,1	-0,10	PASS			
-20	-20,1	-0,10	PASS			
+25	+25,2	0,20	PASS			
+70	+70,1	0,10	PASS			
+85	+85,1	0,10	PASS			
+125	+125,1	0,10	PASS			
+150	+150,2	0,20	PASS			
+170	+170,1	0,10	PASS			

External sensor type K Specification [± 2°C]						
Set Temperature [°C]	As Found Temperature [°C]	As Found Deviation	As Found Result	As Left Temperature [°C]	As Left Deviation	As Left Result
-40	-39,9	0,10	PASS			
-20	-19,8	0,20	PASS			
+25	+24,8	-0,20	PASS			
+70	+70,1	0,10	PASS			
+85	+84,9	-0,10	PASS			
+125	+124,8	-0,20	PASS			
+150	+150,1	0,10	PASS			
+170	+170,1	0,10	PASS			

External sensor type T Specification [± 2°C]						
Set Temperature [°C]	As Found Temperature	As Found Deviation	As Found Result	As Left Temperature	As Left Deviation	As Left Result
-40	-39,8	0,20	PASS			
-20	-19,8	0,20	PASS			
+25	+24,8	-0,20	PASS			
+70	+69,8	-0,20	PASS			
+85	+84,8	-0,20	PASS			
+125	+129,8	4,80	FAIL			
+150	+150,0	0,00	PASS			
+170	+169,9	-0,10	PASS			

If "As Found" Calibration **OK**, then only values in "As Found"

If optimized, then values in "As Found" and "As Left" -> **Remark required**

If adjusted (As Found FAIL), then values in "As Found" and "As Left" -> **Remark required**